



505 - 6th Street, Suite 200
New Westminster, BC V3L 0E1

Toll Free: 1-866-566-SAFE
Fax: (778) 396 - 2064
www.safetyauthority.ca

ALBERTA BOILERS SAFETY ASSOCIATION
9410 20 AVENUE NW
EDMONTON AB T6N 0A4

Date: June 7, 2013
Account #: 34986
Journal #: 56818
Our File #: 5440409

Attn: CYNTHIA FORMANIUK

Re: Application for Design Registration

The design, as detailed in your, ABSA Tracking No 2006-06007 & 2013-01878, for a Fitting is accepted for registration as follows:

Registered To: ROSEMOUNT ANALYTICAL INC. **CRN:** 0F07880.21

MAWT: 212 deg F

MAWP: 250 psig

Drawing #: PDS 71-398

Drawing Revision: E

Conditions Of Registration:

1- TUPH Sensors, models 398 and 398R

2- MDMT:

(a). -20 deg F. for models using carbon steel fixtures.

(b). -50 deg F. for models using stainless steel fixtures.

3- Registration Expiry Date; August 30 ,2016 with the condition of a valid manufacturer's Quality Management System Certificate of Registration.

Contact me if you have any questions. The invoice for registration will be forwarded under separate cover.

ALIREZA BONAKDAR
(778) 396-2032
Alireza.Bonakdar@safetyauthority.ca
Design Engineer

cc: