

August 08, 2013

Attention: Beth Covey
EMERSON PROCESS MANAGEMENT
ROSEMOUNT ANALYTICAL
2400 BARANCA PKWY
IRVINE, CA 92606

The design submission, tracking number 2013-03429, originally received on May 07, 2013 was surveyed and accepted for registration as follows:

CRN : 0F13670.2 **Accepted on:** August 08, 2013

Reg Type: New Design **Expiry Date:** August 08, 2023

Drawing No. : 1700754 SHT.3-4 Rev A As Noted

Fitting type: MODEL 226 TOROIDAL CONDUCTIVITY SENSOR

Design registered in the name of : EMERSON PROCESS MANAGEMENT

The registration is conditional on your compliance with the following notes:

The backbone material is A351 CF8/CF8A as per the manufacturer confirmation.


Model 226 is rated to 295PSI at 0/120 deg.C.

This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form. This registration is valid only until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

Enclosed are stamped prints for your reference.

Sincerely,



ONSHCHENKO, TETYANA, P. Eng.



the pressure equipment safety authority

STATUTORY DECLARATION
Registration of Fittings

In this space, show facsimile of manufacturer's logo or trademark as it will appear on the fitting.



I, Dana Crowley,

Manager of Compliance and Test Engineering

(company title, e.g. vice president, plant manager, chief engineer) (must be in a position of authority)

of Emerson Process Management, Rosemount Analytical

(name of manufacturer)

located at 2400 Barranca Pkwy, Irvine, CA and Circuito del Progreso #27, Parque Industrial Progreso, Mexicali CP-21190

(plant address)

do solemnly declare that the fittings listed hereunder, which are subject to the Safety Codes Act (check one)

[] comply with the requirements of [] which specifies the dimensions, (title of recognized North American Standard)

materials of construction, pressure/temperature ratings and identification marking of the fittings, or

[X] are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with Product Data Sheet as supported by the attached data which identifies the dimensions, materials of construction, pressure/temperature ratings and the basis for such ratings, and the marking of the fittings for identification.

I further declare that the manufacture of these fittings is controlled by a quality control program which has been verified by the following authority, Det Norske Veritas (DNV) as being suitable for the manufacture of these fittings to the stated standard. The fittings covered by this declaration, for which I seek registration, are category F type fittings


In support of this application, the following information, calculations and/or test data are attached:

226 ASME Compliance Drawing PN 1700754; proof pressure test report RAI 03, rev.1; and 226 Product Data Sheet PDS 71-226/228/revG

DECLARED before me at Irvine in the County of Orange

this 6th day of May, 2013 (Month) (Year)

(print) CATHERINE ANN STIP (sign) Catherine Ann Stip (A Commissioner for Oaths)



(Signature of Applicant) Dana Crowley

For Office Use Only

To the best of my knowledge and belief, the application meets the requirements of the Safety Codes Act and CSA Standard B51, Clause 4.2, and is accepted for registration in Category F

Registration Number: OF13670.2

Date Registered: AUG 08 2013

(For the Administrator/Chief Inspector of Alberta)

Expiry Date: August 8, 2013