

November 08, 2012

**Attention:** Beth Covey  
EMERSON PROCESS MANAGEMENT  
PO BOX 270658  
ST LOUIS, MO

The design submission, tracking number 2012-05374, originally received on August 02, 2012 was surveyed and accepted for registration as follows:

**CRN :** 0F12814.2 **Accepted on:** November 08, 2012

**Reg Type:** New Design **Expiry Date:** November 08, 2022

**Drawing No. :** 1700749 Rev A

**Fitting type:** 3300/VP, 3400/VP & 3500/VP PH SENSORS

Design registered in the name of : ROSEMOUNT ANALYTICAL INC

**The registration is conditional on your compliance with the following notes:**

*The pressure and temperature ratings are as indicated by the drawing:*

- Models 3300HT/HTVP and 3400HT/HTVP rated 400 psi @ +5 / 155 C;
- Models 3500/VP rated 100 psi @ +5 / 120 C;

*This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form. This registration is valid only until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.*

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3347 or fax (780) 437-7787 or e-mail Preda@absa.ca.

Sincerely,



PREDA, DRAGOS, P. Eng.